

# SAMPLING LINE FOR DUST IN EMISSION



**NOTE:** The isokinetic probes are equipped with a filter holder but, if it is necessary, they must be completed with the basket.  
For temperatures higher than 500 ° C is recommended the addition of the serpentine cod. A.3041 between sampling probe and